

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 16 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130404</u>
Name	<u>John W Carroll</u>	ID# <u>6775</u> Date <u>03-14-2010</u>
Agency	<u>Fort Greely Police Dept. BIG4</u>	Phone # <u>907-873-9134</u>
Instrument Location	<u>Ft. Greely Police Bookings Rm Bldg 663</u>	
Alco S/N	<u>X172958</u>	Target Value <u>.087</u> High Pressure <u><del>800</del> 800</u>
Alco Test Values	<u>.086</u> 1 <sup>st</sup> Alco	<u>.085</u> 2 <sup>nd</sup> Alco
Signature	<u>John W Carroll</u>	
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15<sup>th</sup> day of April, 2010.

Carolyn M. Noland  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

BT9 03/00



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MARCH 14, 2010

OPERATOR'S NAME:

CARROLL, JOHN

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 000000

DEPT/AGENCY: FT GREELY PD B164

CASE/REPORT: 00-0000

TEST TYPE: V

ALCO TARGET VALUE: .087 ✓

ALCO S/N: X172958 ✓

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 28.53 in  
ALCO TARGET .082 22:23  
BLANK TEST .000 22:24  
INTERNAL STANDARD VERIFIED 22:24  
ALCO TV 28.53 in .086 ✓ 22:24  
BLANK TEST .000 22:25  
SUBJECT SAMPLE .000 22:26  
BLANK TEST .000 22:27  
ALCO TV 28.53 in .085 ✓ 22:27  
BLANK TEST .000 22:28

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MARCH 14, 2010

TIME 22:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 28.53 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~